

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-07-01 16:50:10

Date of Next Testing: 2028-06-30 16:50:10

Product Name: **TOC Standards**

CAS Number:

Specifications & Purity: Description: Total Organic Carbon Standard 50ppm

Lot #: I2413653

SKU #: **T197153**

Version: **3**

Analysis #: **727331**

Storage Temperature: Store at 2-8°C, Protected from light

Parameter	Limit Values	Result
Assay	49-51 (ppm)	49.98000 ppm
Related expanded uncertainty(k=2)	0-5 (%)	4.98000 %
Solvent(Water)	Conform	Conform



John Su
QA & QC Manager