



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-07-21 10:54:16
Date of Next Testing: 2027-07-21 10:54:16
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description:Ratio 0.2 NTU
Lot #: L2129375
SKU #: **T197167**
Version: **1**
Analysis #: **543628**

Parameter	Limit Values	Result
Assay	0.18-0.22(NTU)	0.18900 NTU

John Su
QA & QC Manager