



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2022-04-13 14:31:51
Date of Next Testing: 2024-04-12 14:31:51
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description:Ratio 0.5 NTU
Lot #: E2011090
SKU #: **T197169**
Version: **0**
Analysis #: **086650**
Storage Temperature: Store at 2-8°C

Parameter	Limit Values	Result
Assay	0.45-0.55 (NTU)	0.46000 NTU

John Su
QA & QC Manager