



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-12-21 10:10:17
Date of Next Testing: 2025-12-20 10:10:17
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description:Ratio 5 NTU
Lot #: G1908100
SKU #: **T197173**
Version: **0**
Analysis #: **307148**
Storage Temperature: Room temperature

Parameter	Limit Values	Result
Assay limits	4.75-5.25 (NTU)	4.77000 NTU

John Su
QA & QC Manager