



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-07-21 11:00:30
Date of Next Testing: 2027-07-21 11:00:30
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description:Ratio 20 NTU
Lot #: L2129378
SKU #: **T197175**
Version: **1**
Analysis #: **543631**

Parameter	Limit Values	Result
Assay	19-21(NTU)	20.08100 NTU

John Su
QA & QC Manager