



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-07-21 10:51:45
Date of Next Testing: 2027-07-21 10:51:45
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description:Ratio 40 NTU
Lot #: L2129367
SKU #: **T197176**
Version: **1**
Analysis #: **543623**

Parameter	Limit Values	Result
Assay	38-42(NTU)	40.10000 NTU

John Su
QA & QC Manager