



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-07-21 10:52:14
Date of Next Testing: 2027-07-21 10:52:14
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description:Ratio 50 NTU
Lot #: L2129370
SKU #: **T197177**
Version: **1**
Analysis #: **543624**

Parameter	Limit Values	Result
Assay	47.5-52.5(NTU)	50.20000 NTU

John Su
QA & QC Manager