



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-07-21 10:52:46
Date of Next Testing: 2027-07-21 10:52:46
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description:Ratio 60 NTUv
Lot #: L2129371
SKU #: **T197178**
Version: **1**
Analysis #: **543625**

Parameter	Limit Values	Result
Assay	57-63(NTU)	61.45000 NTU

John Su
QA & QC Manager