



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-08-19 10:59:27
Date of Next Testing: 2027-08-19 10:59:27
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description:Ratio 90 NTU
Lot #: B2209326
SKU #: **T197179**
Version: **1**
Analysis #: **555408**

Parameter	Limit Values	Result
Assay	85.5-94.5(NTU)	92.22800 NTU

John Su
QA & QC Manager