



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-07-21 13:17:57
Date of Next Testing: 2027-07-21 13:17:57
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description: Non Ratio 5 NTU
Lot #: A2205613
SKU #: **T197195**
Version: **1**
Analysis #: **543672**

Parameter	Limit Values	Result
Turbidity	4.75-5.25 (NTU)	5.10000 NTU

John Su
QA & QC Manager