



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-06-12 11:45:38
Date of Next Testing: 2028-06-11 11:45:38
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description: Non Ratio 5 NTU
Lot #: L2205539
SKU #: **T197195**
Version: **1**
Analysis #: **720374**
Storage Temperature: Room temperature

Parameter	Limit Values	Result
Turbidity	4.75-5.25 (NTU)	4.99800 NTU

John Su
QA & QC Manager