



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-05-19 10:16:46
Date of Next Testing: 2028-05-18 10:16:46
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description: Non Ratio 20 NTU
Lot #: G2431564
SKU #: **T197197**
Version: **1**
Analysis #: **704404**
Storage Temperature: Room temperature

Parameter	Limit Values	Result
Turbidity	19-21 (NTU)	19.98000 NTU

John Su
QA & QC Manager