



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-07-21 13:17:31
Date of Next Testing: 2027-07-21 13:17:31
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description: Non Ratio 50 NTU
Lot #: A2205612
SKU #: **T197199**
Version: **1**
Analysis #: **543671**

Parameter	Limit Values	Result
Turbidity	47.5-52.5 (NTU)	51.01000 NTU

John Su
QA & QC Manager