



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-10-12 14:24:00
Date of Next Testing: 2025-10-11 14:24:00
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description: Non Ratio 60 NTU
Lot #: A2205617
SKU #: **T197200**
Analysis: **284504**

Parameter	Limit Values	Result
Turbidity	57 - 63 (NTU)	59.00000 NTU

John Su
QA & QC Manager