



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-03-02 16:27:46
Date of Next Testing: 2028-03-01 16:27:46
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description:Non Ratio 60 NTU
Lot #: B2628705
SKU #: **T197200**
Version: **2**
Analysis #: **660491**
Storage Temperature: Room temperature

Parameter	Limit Values	Result
Turbidity	57 - 63 (NTU)	60.00000 NTU
Related expanded uncertainty(k=2)	0-5 (%)	5.00000 %

John Su
QA & QC Manager