



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2024-05-27 14:40:59
Date of Next Testing: 2026-05-27 14:40:59
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description: Non Ratio 60 NTU
Lot #: E2420371
SKU #: **T197200**
Version: **1**
Analysis #: **345085**

Parameter	Limit Values	Result
Turbidity	57 - 63 (NTU)	57.00000 NTU

John Su
QA & QC Manager