



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-06-19 18:20:35
Date of Next Testing: 2027-06-19 18:20:35
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description: Non Ratio 60 NTU
Lot #: F2518550
SKU #: **T197200**
Version: **1**
Analysis #: **528568**

Parameter	Limit Values	Result
Turbidity	57 - 63 (NTU)	60.50000 NTU

John Su
QA & QC Manager