



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-08-19 11:16:00
Date of Next Testing: 2027-08-19 11:16:00
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description: Non Ratio 90 NTU
Lot #: B2209376
SKU #: **T197201**
Version: **1**
Analysis #: **555417**

Parameter	Limit Values	Result
Turbidity	85.5-94.5 (NTU)	92.55800 NTU

John Su
QA & QC Manager