



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-10-21 09:49:41
Date of Next Testing: 2027-10-21 09:49:41
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description:Non Ratio 100 NTU
Lot #: A2415585
SKU # **T197202**
Version **1**
Analysis #: **588924**

Parameter	Limit Values	Result
Turbidity	95-105 (NTU)	99.00000 NTU

John Su
QA & QC Manager