



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

## Certificate of Analysis

Date of Testing: 2025-04-11 14:48:05  
Date of Next Testing: 2027-04-11 14:48:05  
Product Name: **Turbidity Standards**  
CAS Number:  
Specifications & Purity: Description: Non Ratio 400 NTU  
Lot #: G23151183  
SKU #: **T197204**  
Version: **1**  
Analysis #: **493595**

| Parameter | Limit Values  | Result        |
|-----------|---------------|---------------|
| Turbidity | 380-420 (NTU) | 403.00000 NTU |

John Su  
QA & QC Manager