



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-07-21 13:18:56
Date of Next Testing: 2027-07-21 13:18:56
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description: Non Ratio 500 NTU
Lot #: A2205618
SKU #: **T197207**
Version: **1**
Analysis #: **543674**

Parameter	Limit Values	Result
Turbidity	475-525 (NTU)	501.00000 NTU

John Su
QA & QC Manager