



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-07-21 13:21:33
Date of Next Testing: 2027-07-21 13:21:33
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description: Non Ratio 800 NTU
Lot #: A2205621
SKU #: **T197209**
Version: **1**
Analysis #: **543677**

Parameter	Limit Values	Result
Turbidity	760-840 (NTU)	793.00000 NTU

John Su
QA & QC Manager