



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-10-21 09:50:24
Date of Next Testing: 2027-10-21 09:50:24
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description:Non Ratio 800 NTU
Lot #: A2415587
SKU # **T197209**
Version **1**
Analysis #: **588925**

Parameter	Limit Values	Result
Turbidity	760-840 (NTU)	816.00000 NTU

John Su
QA & QC Manager