



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2022-06-29 11:06:22
Date of Next Testing: 2024-06-28 11:06:22
Product Name: **Turbidity Standards**
CAS Number:
Specifications & Purity: Description:Non Ratio 4000 NTU
Lot #: F2225532
SKU #: **T197211**
Version: **2**
Analysis #: **095681**
Storage Temperature: Store at 2-8°C, Protected from light

Parameter	Limit Values	Result
Assay limits	3920-4080 (NTU)	4000.00000 NTU

John Su
QA & QC Manager