



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

## Certificate of Analysis

Date of Testing: 2026-08-04 14:15:53  
Date of Next Testing: 2028-08-03 14:15:53  
Product Name: **Turbidity Standards**  
CAS Number:  
Specifications & Purity: Description:Non Ratio 4000 NTU  
Lot #: G2629676  
SKU #: **T197211**  
Version: **2**  
Analysis #: **744294**  
Storage Temperature: Store at 2-8°C, Protected from light

| Parameter                         | Limit Values    | Result         |
|-----------------------------------|-----------------|----------------|
| Assay limits                      | 3920-4080 (NTU) | 4000.00000 NTU |
| Related expanded uncertainty(k=2) | 0-4 (%)         | 3.00000 %      |

John Su  
QA & QC Manager