

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-08-27 09:59:21
Date of Next Testing: 2030-08-26 09:59:21
Product Name: **Tris-Hydrochloride Buffer**
CAS Number:
Grades & Purity: 1M , pH 8.0 , sterile
Lot #: H2621002
SKU #: **T301502**
Version: **2**
Analysis #: **762877**
Storage Temperature: Room temperature

Parameter	Limit Values	Result
Concentration	0.95-1.05 (mol/L)	1.00000 mol/L
Conductivity(25°C)(Test by Conductivity Meter)	29.901-34.194 (mS/cm)	31.84300 mS/cm
PH (25°C)	7.9-8.1	7.99300
Appearance (T301502)	Colorless, transparent liquid	Colorless, transparent liquid



John Su
QA & QC Manager