

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-06-12 14:00:39

Date of Next Testing: 2029-06-11 14:00:39

Product Name: **Tris-Hydrochloride Buffer**

CAS Number:

Specifications & Purity: 0.5M , pH 6.8 , sterile

Lot #: F2116004

SKU # **T301507**

Version **0**

Analysis #: **528217**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
Concentration	0.48-0.52 (mol/L)	0.50000 mol/L
pH(25 °C)	6.7-6.9	6.80000
Appearance (T301507)	Colorless, transparent liquid	Colorless, transparent liquid



John Su
QA & QC Manager