

Certificate of Analysis

Date of Testing: 2023-04-10 11:23:46
Date of Next Testing: 2025-04-09 11:23:46
Product Name: **TE buffer**
CAS Number:
Specifications & Purity: 10× pH 7.0
Lot #: C2117059
SKU #: **T301523**
Analysis: **208274**

Parameter	Limit Values	Result
Concentration	0.0098-0.012 (mol/L)	0.01 mol/L
PH(25 °C)	6.8-7.2	7.00
Appearance (T301523)	Colorless, transparent liquid	Colorless, transparent liquid



John Su
QA & QC Manager