

Certificate of Analysis

Date of Testing: 2024-07-01 17:18:13
Date of Next Testing: 2026-07-01 17:18:13
Product Name: **TE buffer**
CAS Number:
Specifications & Purity: 20× pH 8.0
Lot #: H2201117
SKU #: **T301540**
Version: **2**
Analysis #: **378742**

| Parameter | Limit Values | Result |
|----------------------|-------------------------------|-------------------------------|
| pH(25°C , 1x) | 7.9-8.1 | 8.00000 |
| Appearance (T301540) | Colorless, transparent liquid | Colorless, transparent liquid |
| Sterile | Pass | Pass |



John Su
QA & QC Manager