

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-03-27 15:47:58

Date of Next Testing: 2027-03-27 15:47:58

Product Name: **Urine Bilirubin Qualitative Detection Kit (Harrison's Method)**

CAS Number:

Specifications & Purity:

Lot #:

SKU #

Version:

Analysis #:

BioReagent

ZJ26F0333460

U1509474-CG50T
1

028991

Parameter	Limit Value	Result
Seal Integrity Testing	Conform	Conform
Appearance (U1509474A)	Colorless and transparent solution, free of sediment	Colorless and transparent solution, free of sediment
Odor (U1509474A)	Odorless	Odorless
Density (U1509474A) (1.02 - 1.12 g/mL)	Conform	Conform
Appearance (U1509474B)	Light-yellow and transparent solution, free of sediment	Light-yellow and transparent solution, free of sediment
Odor (U1509474B)	Pungent odor	Pungent odor
Density (U1509474B) (1.08 - 1.14 g/mL)	Conform	Conform
Performance	Conform	Conform



Sophia Cheng
Quality Assurance