

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-12-22 15:00:11

Date of Next Testing: 2025-12-21 15:00:11

Product Name: **Viscosity Standards expressed in cSt**

CAS Number:

Specifications & Purity: 20°C:0.47;25°C:0.45;37.78°C:0.41;40°C:0.40;50°C

Lot #: C2214280

SKU #: **V197297**

Version: **2**

Analysis #: **307965**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
Viscosity(20°C)	0.4828-0.4842(mm ² /s)	0.48350 mm ² /s
Viscosity(25°C)	0.4623-0.4635(mm ² /s)	0.46290 mm ² /s
Viscosity(37.78°C)	0.4157-0.4169(mm ² /s)	0.41630 mm ² /s
Viscosity(40°C)	0.4085-0.4097(mm ² /s)	0.40910 mm ² /s



John Su
QA & QC Manager