

ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-04-18 14:06:31

Date of Next Testing: 2025-04-17 14:06:31

Product Name: **Viscosity Standards expressed in cSt**

CAS Number:

Specifications & Purity: 20°C:19000;25°C:12000;37.78°C:4000;40°C:3400;50°C:1700

Lot #: G2113272

SKU #: **V298527**

Version: **3**

Analysis #: **213048**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
Viscosity(20°C)	15855-15946 (mPa·S)	15900.00000 mPa·S
Viscosity(25°C)	10050-10124 (mPa·S)	10100.00000 mPa·S
viscosity(40°C)	2935-2954 (mPa·S)	2940.00000 mPa·S
viscosity(50°C)	1457-1466 (mPa·S)	1460.00000 mPa·S
Viscosity(37.78°C)	3474-3496(mPa.s)	3482.00000 mPa·S



John Su
QA & QC Manager