



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2025-09-16 10:40:01

Date of Next Testing: 2027-09-16 10:40:01

Product Name: **Viscosity Standards expressed in cSt**

CAS Number:

Specifications & Purity: 20°C:19000;25°C:12000;37.78°C:4000;40°C:3400;50°C:1700

Lot #: I2515065

SKU #: **V298527**

Version **3**

Analysis #: **514642**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
Viscosity(20°C)	17000-19000 (mm ² /s)	17120.00000 mm ² /s
Viscosity(25°C)	10000-12000 (mm ² /s)	10920.00000 mm ² /s
Viscosity(40°C)	3200-3400(mm ² /s)	3242.00000 mm ² /s
Viscosity(50°C)	1600-1700 (mm ² /s)	1626.00000 mm ² /s
Viscosity(37.78°C)	3800-4000(mm ² /s)	3827.00000 mm ² /s

John Su
QA & QC Manager