

Certificate of Analysis

Date of Testing: 2023-08-29 12:02:46

Date of Next Testing: 2024-08-28 12:02:46

Product Name: **Water**

CAS Number: 7732-18-5

Specifications & Purity: Used for cell culture and molecular biology, sterilization filtration

Lot #: H2329645

SKU #: **W119421**

Analysis: **269217**

Parameter	Limit Values	Result
DNases	0-0.004 (mg/L)	0.00 mg/L
RNases	0-10 (mg/L)	2.40 mg/L
PH	5-7	6.50
Appearance (W119421)	Colorless Liquid	Colorless liquid
Bacterial	< 1cfu/μl	< 1cfu/μl
Cell Culture Test	pass	Pass
Cell Line	Cell Line - Cell Types	Cell Line - Cell Types
Endotoxin Level	< 0.001 EU/mL	< 0.001 EU/mL
Sterility	Pass	Pass



John Su
QA & QC Manager