



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2023-08-01 13:38:02
Date of Next Testing: 2025-07-31 13:38:02
Product Name: **Turbidity Standard**
CAS Number:
Specifications & Purity: 400 NTU in water
Lot #: J2129209
SKU #: **W119511**
Analysis: **258051**
Storage Temperature: Store at 2-8°C

Parameter	Limit Values	Result
Turbidity	388-412 (FTU)	400.00 FTU

John Su
QA & QC Manager