



ALADDIN SCIENTIFIC CORPORATION

14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2026-08-04 16:18:46

Date of Next Testing: 2027-08-04 16:18:46

Product Name: **Water**

CAS Number: 7732-18-5

Specifications & Purity: Low Endotoxin

Lot #: G2624255

SKU #: **W1492513**

Version: **2**

Analysis #: **754857**

Storage Temperature: Room temperature

Parameter	Limit Values	Result
Endotoxin	0-0.25 (EU/ml)	0.12750 EU/ml
(NO3)Nitrate	0-0.2 (ppm)	0.12700 ppm
Osmolality	0-20 (mOsmol/kg)	1.54000 mOsmol/kg
pH	4.0-7.5	6.50000
Total organic carbon	0-500 (ppb)	1.56000 ppb
Appearance(W1492513)	Colorless Liquid	Colorless liquid
Conductivity EP/USP-Point of fill	Pass	Pass
Sterility(Negative)	pass	Pass

John Su
QA & QC Manager