

ALADDIN SCIENTIFIC CORPORATION  
14078 Meridian Parkway, Riverside, CA. 92518

### Certificate of Analysis

Date of Testing: 2026-01-06 17:42:25  
Date of Next Testing: 2029-01-05 17:42:25  
Product Name: **Water**  
CAS Number: 7732-18-5  
Specifications & Purity: UltraBio™, Ultra pure, Embryo cell culture tested, for embryo transfer, sterile-filtered  
Lot #: A2606636  
SKU #: **W433893**  
Version: **1**  
Analysis #: **635875**

Parameter	Limit Values	Result
pH	5.0-7.0	6.50
Endotoxin Level	0-1 (EU/ml)	0.02500 EU/ml
Appearance(W433893)	Clear Liquid	Clear liquid
Sterility	Pass	Pass
Embryo Transfer	Pass	Pass



John Su  
QA & QC Manager