

ALADDIN SCIENTIFIC CORPORATION
14078 Meridian Parkway, Riverside, CA. 92518

Certificate of Analysis

Date of Testing: 2024-07-18 09:52:03
Date of Next Testing: 2027-07-18 09:52:03
Product Name: **Water**
CAS Number: 7732-18-5
Specifications & Purity: for embryo transfer, sterile-filtered, BioReagent Plus, suitable for mouse embryo cell culture
Lot #: G2417289
SKU #: **W433893**
Version: **1**
Analysis #: **389657**

| Parameter | Limit Values | Result |
|---------------------|--------------|---------------|
| PH | 5.0-7.0 | 6.50 |
| Endotoxin Level | 0-1 (EU/ml) | 0.02700 EU/ml |
| Appearance(W433893) | Clear Liquid | Clear liquid |
| Sterility | Pass | Pass |
| Embryo Transfer | Pass | Pass |



John Su
QA & QC Manager