

Shanghai Aladdin Biochemical Technology Co., LTD
809 Chuhua Branch Road, Fengxian District, Shanghai 201400, China

Certificate of Analysis

Date of Testing: 2025-01-04 15:02:00
Date of Next Testing: 2027-01-04 15:02:00
Product Name: **Membrane Filter**
CAS Number:
Specifications & Purity: i-Quip™, Membrane Dia.:47mm
Lot #: K2430008
SKU #: **M2637**
Version: **1**
Analysis #: **432897**

Parameter	Limit Values	Result
Appearance(Smooth and flat, without burrs)	pass	Pass
material: Hydrophobic PTFE	Pass	Pass
Housing Material: PP	Pass	Pass
Diameter(47±2.4mm)	pass	Pass
Pore diameter(0.22±0.01μm)	Pass	Pass
Thickness:130-150μm	Pass	Pass



John Su
QA & QC Manager